

IonTamer™ PS Iong-term stability

CONFIDENTIAL

2023-01-24 | Sébastien Gasc

2023-01-24 CONFIDENTIAL

Long-term stability

Measurement description

IonTamer PS-H operation

- Continuously ON
- High Emission (2 mA, 70 eV)

Long-term test conditions

- Continuous Cl_2 inlet @ $1.7 \cdot 10^{-7}$ mbar (= $1.257 \cdot 10^{-7}$ Torr)
- Total test duration: 440 h of cumulated operation with Cl₂ inlet

Reference measurements

- Conducted every 2-3 days, right before closing the Cl₂ inlet or right after opening the Cl₂ inlet
- Gas inlet @ 1.7 10⁻⁷ mbar (= 1.257 10⁻⁷ Torr): mixture of Cl₂ & noble gas mixture (1 ‰ krypton, rest. argon) in 5:1 proportion
- 15-minute integration time

Data analysis considerations

- Data have been corrected for the specific sensitivity of the lonTamer to chlorine, argon, and krypton
- Theoretical value of 1 ‰ Kr in Ar has been corrected (to 0.572 ‰) to take into account isotopic ratios for ⁴⁰Ar and ⁸⁴Kr

Results


| Time (UTC) | ⁸⁴ Kr/ ⁴⁰ Ar ratio [‰] |
|------------------|---|
| 2022-12-01 15:42 | 0.549 |
| 2022-12-02 17:26 | 0.579 |
| 2022-12-03 17:33 | 0.538 |
| 2022-12-04 11:14 | 0.559 |
| 2022-12-09 13:26 | 0.652 |
| 2022-12-12 09:58 | 0.606 |
| 2022-12-14 10:39 | 0.573 |
| 2022-12-16 14:06 | 0.536 |
| 2022-12-19 09:33 | 0.509 |
| 2022-12-22 09:26 | 0.551 |





Long-term Cl₂ measurement

Variation of the ⁸⁴Kr/⁴⁰Ar ratio over time during long-term Cl₂ measurements



The lonTamer[™] reliably measured the partial pressure of Cl₂ over more than 440 h of cumulated operation. Variations in the signal are due to variations in the process chamber the lonTamer[™] was connected to.

While measuring in a harsh environment with corrosive gas (Cl₂), the instrument remained quantitative and was capable to measure a stable and correct Kr/Ar ratio.

Reference measurements were conducted right before closing the Cl₂ inlet or right after opening the Cl₂ inlet.